

Certificate FI20/966870

Hangzhou Shanyou Medical Equipment Co., Ltd.

NO. 138 Louta Development Zone, Guancun Village, Louta Town, Xiaoshan District, Hangzhou,Zhejiang 311266, China

It is certified that the manufacturer's technical file and the PPE product detailed on page 2 have been assessed and found to be in accordance with


Regulation (EU) 2016/425

Module B, EU type-examination

This certificate is valid from 12 October 2020 until 12 October 2025

1. Certified since 12 October 2020

Authorised by



FINAS
Finnish Accreditation Service
S003 (EN ISO/IEC 17065)

SGS FIMKO OY, Notified Body 0598

Takomotie 8, FI-00380, Helsinki, Finland t +358 9 696 361 f +358 9 692 5474 www.sgs.com

Page 1 of 2





This document is issued by the Company subject to its General Conditions of Certification Services accessible at www.sgs.com/terms_and_conditions.htm.
Attention is drawn to the limitations of liability, indemnification and jurisdictional issues established therein. The authenticity of this document may be verified at http://www.sgs.com/en/certified-clients-and-products/certified-client-directory.

Any unauthorized alleration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be presected to the fullest



Certificate FI20/966870, continued

Hangzhou Shanyou Medical Equipment Co., Ltd.

Regulation (EU) 2016/425

Module B, EU type-examination

Issue 1

PPE Product

WORK (logo) SYMS-1 particle filtering half mask, consisting of a four layer (Polypropylene/Hot air cotton) disposable face mask, with Polyethylene covered Iron wire nose clip and Polyamide/Spandex mask straps.

It is certified that the manufacturer's technical file and the above mentioned PPE have been assessed and found to meet the applicable Essential Health and Safety Requirements in Annex II of Regulation (EU) 2016/425 Personal Protective Equipment

The following have been applied:

EN 149: 2001 +A1:2009 (Respiratory protective devices- filtering half masks to protect against particles) device classification: FFP2 NR.

This certificate is issued on strict condition that appropriate checks on manufactured PPE, as detailed in Article 19(c) of the regulation are implemented and maintained while the model is in production.

Certification is based on technical file reference:

WK01, Dated 22.07.2020

SGS Reference Number UK/CRS 242026

This certificate remains the property of SGS Fimko Oy to whom it must be returned on request

